




Potvrda o akreditaciji Accreditation Certificate

Ovime se utvrđuje da je
This is to recognize that

Zavod za javno zdravstvo Požeško-slavonske županije
Služba za zdravstvenu ekologiju
Županijska 9, HR-34000 Požega

osposobljen prema zahtjevima norme
is competent according to
HRN EN ISO/IEC 17025:2017
(ISO/IEC 17025:2017;
EN ISO/IEC 17025:2017)
za/to carry out

Ispitivanje i uzorkovanje voda
Ispitivanje meda
Testing and sampling of waters
Testing of honey

u području opisanom u prilogu koji je sastavni dio ove potvrde o akreditaciji.
for the scope described in the annex which is the constituent part of this accreditation certificate.

Br./No.: 1239
Klasa/Ref.No.: 383-02/24-30/029
Urbroj/Id.No.: 569-02/3-25-20
Zagreb, 2025-04-18

Akreditacija istječe-Accreditation expiry: 2030-04-17
Prva akreditacija-Initial accreditation: 2010-04-28

HAA je potpisnica multilateralnog sporazuma s Europskom organizacijom za akreditaciju (EA)
HAA is a signatory of the European co-operation for Accreditation (EA) Multilateral Agreement

Ravnateljica:
Director General:
mr. sc. Mirela Zečević



Hrvatska akreditacijska agencija
Croatian Accreditation Agency

PRIOLOG POTVRDI O AKREDITACIJI br: 1239

Annex to Accreditation Certificate Number:

Klasa/Ref. No.: 383-02/24-30/029

Urbroj/Id. No.: 569-02/3-25-19

Datum izdanja priloga /Annex Issued on: 2025-04-18

Zamjenjuje prilog/Replaces Annex:

Klasa/Ref. No.: 383-02/19-30/031

Urbroj/Id. No.: 569-02/3-23-40

Datum/Date: 2023-10-03

Norma: HRN EN ISO/IEC 17025:2017

Standard: (ISO/IEC 17025:2017; EN ISO/IEC 17025:2017)

Akreditacija istječe: 2030-04-17

Accreditation expiry:

Prva akreditacija: 2010-04-28

Initial accreditation:

Akreditirani laboratorij

Accredited Laboratory

Zavod za javno zdravstvo Požeško-slavonske županije

Služba za zdravstvenu ekologiju

Županijska 9, HR-34000 Požega

Područje akreditacije:

Scope of Accreditation:

Ispitivanje i uzorkovanje voda

Ispitivanje meda

Testing and sampling of waters

Testing of honey

Važeće izdanje Priloga dostupno je na web adresi: www.akreditacija.hr /
Valid issue of the Annex is available at the web address: www.akreditacija.hr

Ravnateljica:

Director General:

mr. sc. Mirela Zečević

PODRUČJE AKREDITACIJE / SCOPE OF ACCREDITATION

Br. No.	Materijali/Proizvodi Materials/Products	Vrsta ispitivanja/Svojstvo Type of test/Property <i>Raspon/Range</i>	Metoda ispitivanja Test method
1.	Voda za ljudsku potrošnju, prirodne izvorske, prirodne mineralne i stolne, bazenske vode <i>Water for human consumption, natural spring water, natural mineral water and table water, pool water</i>	Određivanje električne vodljivosti <i>Determination of electrical conductivity</i> (od/from 1 do/to 2000) $\mu\text{S/cm}$	HRN EN 27888:2008 (ISO 7888:1985; EN 27888:1993)
2.		Određivanje pH vrijednosti <i>Determination of pH value</i> (od/from 5,0 do/to 10,0) pH	HRN EN ISO 10523:2012 (ISO 10523:2008; EN ISO 10523:2012)
3.		Određivanje mutnoće <i>Determination of turbidity</i> (od/from 0,05 do/to 200) NTU	HRN EN ISO 7027-1:2016 (ISO 7027-1:2016; EN ISO 7027-1:2016)
4.		Brojanje <i>Escherichia coli</i> i koliformnih bakterija – Metoda membranske filtracije <i>Enumeration of Escherichia coli and coliform bacteria – Membrane filtration method</i>	HRN EN ISO 9308-1:2014 (ISO 9308-1:2014; EN ISO 9308-1:2014) HRN EN ISO 9308-1:2014/A1:2017 (ISO 9308-1:2014/ Amd 1:2016; EN ISO 9308-1:2014/A1:2017)
5.		Detekcija i brojenje enterokoka – Metoda membranske filtracije <i>Detection and enumeration of enterococci – Membrane filtration method</i>	HRN EN ISO 7899-2:2000 (ISO 7899-2:2000, EN ISO 7899-2:2000)
6.		Detekcija i brojenje <i>Pseudomonas aeruginosa</i> – Metoda membranske filtracije <i>Detection and enumeration of Pseudomonas aeruginosa – Methode by membrane filtration</i>	HRN EN ISO 16266:2008 (ISO 16266:2006; EN ISO 16266:2008)
7.		Brojenje uzgojenih mikroorganizama – broj kolonija naciepljivanem na hranjivi agar <i>Enumeration of culturable microorganisms-colony count by inoculation in a nutrient agar culture medium</i>	HRN EN ISO 6222:2000 (ISO 6222:1999, EN ISO 6222:1999)

Br. No.	Materijali/Proizvodi Materials/Products	Vrsta ispitivanja/Svojstvo Type of test/Property Raspon/Range	Metoda ispitivanja Test method
8.	Voda za ljudsku potrošnju Water for human consumption	Uzorkovanje Sampling	HRN ISO 5667-5:2011 (ISO 5667-5:2006) HRN EN ISO 19458:2008 (ISO 19458:2006; EN ISO 19458:2006)
9.	Bazenska voda Pool water	Uzorkovanje Sampling	HRN EN ISO 19458:2008 (ISO 19458:2006; EN ISO 19458:2006)
10.	Otpadne vode Waste water	Uzorkovanje Sampling	HRN ISO 5667-10:2020 (ISO 5667-10:2020)
11.		Određivanje nitrata Nitrate determination Granica kvantifikacije/ Quantification limit: 0,23 mg/l N	Vlastita metoda In-house method PL-300 Izdanje/ Edition: 02 Datum/ Date: 1.9.2019. temeljena na/based on Hach-Lang kivetnim testovima Hach-Lange cuvette tests LCK 339 LCK 340
12.	Med Honey	Određivanje vode u medu Determination of water in honey	Vlastita metoda / In-house method PL-315 Izdanje/ Edition: 03 Datum/ Date: 5.4.2023. modificirana/modified Harmonised Methods of the International Honey Commission (2009) Method 1